

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF060319\
 Data File : FF004690.D
 Signal(s) : FID2B.ch
 Acq On : 03 Jun 2019 14:03
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Jun 04 01:23:51 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
 Quant Title :
 QLast Update : Thu May 23 07:21:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.705	5700326	46.681 ug/ml
Target Compounds			
1) N-OCTANE	2.311	7327980	50.794 ug/ml
2) N-DECANE	4.418	7611077	51.977 ug/ml
3) N-DODECANE	6.442	7467420	50.462 ug/ml
4) N-TETRADECANE	8.238	7236820	49.267 ug/ml
5) N-HEXADECANE	9.841	7106511	48.132 ug/ml
6) N-OCTADECANE	11.287	6866781	47.297 ug/ml
7) N-EICOSANE	12.600	6750572	46.611 ug/ml
8) N-DOCOSANE	13.803	6543314	46.306 ug/ml
10) N-TETRACOSANE	14.910	6473305	46.709 ug/ml
11) N-HEXACOSANE	15.935	6497175	47.684 ug/ml
12) N-OCTACOSANE	16.889	6615487	48.657 ug/ml
13) N-TRIACONTANE	17.781	6821953	49.333 ug/ml
14) N-DOTRIACONTANE	18.619	6844244	49.477 ug/ml
15) N-TETRATRIACONTANE	19.409	6901144	50.476 ug/ml
16) N-HEXATRIACONTANE	20.156	6683563	50.768 ug/ml
17) N-OCTATRIACONTANE	20.911	6343048	52.747 ug/ml
18) N-TETRACONTANE	21.842	6465989	52.509 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF060319\
Data File : FF004690.D
Signal(s) : FID2B.ch
Acq On : 03 Jun 2019 14:03
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Jun 04 01:23:51 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
Quant Title :
QLast Update : Thu May 23 07:21:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

