

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF060319\
 Data File : FF004692.D
 Signal(s) : FID2B.ch
 Acq On : 03 Jun 2019 15:02
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Jun 04 01:24:25 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
 Quant Title :
 QLast Update : Thu May 23 07:21:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.705	5700270	46.681 ug/ml
Target Compounds			
2) N-DECANE	4.418	7606836	51.948 ug/ml
3) N-DODECANE	6.442	7454717	50.376 ug/ml
4) N-TETRADECANE	8.238	7216132	49.126 ug/ml
5) N-HEXADECANE	9.842	7081010	47.960 ug/ml
6) N-OCTADECANE	11.287	6846466	47.158 ug/ml
7) N-EICOSANE	12.601	6743951	46.565 ug/ml
8) N-DOCOSANE	13.803	6545159	46.319 ug/ml
10) N-TETRACOSANE	14.910	6457041	46.592 ug/ml
11) N-HEXACOSANE	15.935	6454855	47.374 ug/ml
12) N-OCTACOSANE	16.889	6592733	48.490 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF060319\
Data File : FF004692.D
Signal(s) : FID2B.ch
Acq On : 03 Jun 2019 15:02
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Jun 04 01:24:25 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
Quant Title :
QLast Update : Thu May 23 07:21:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

