

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF060622\
 Data File : FF010703.D
 Signal(s) : FID2B.ch
 Acq On : 06 Jun 2022 17:30
 Operator : AJ\MA
 Sample : 10 TRPH STD
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 10 TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/07/2022
 Supervised By :Ankita Jodhani 06/07/2022

Integration File: autoint1.e
 Quant Time: Jun 07 03:22:35 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF060622.M
 Quant Title :
 QLast Update : Tue Jun 07 03:21:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.822	1166573	11.014 ug/ml
Target Compounds			
1) N-OCTANE	1.880	1198327	10.850 ug/ml
2) N-DECANE	4.393	1223798	10.777 ug/ml
3) N-DODECANE	6.559	1242907	10.708 ug/ml
4) N-TETRADECANE	8.381	1264054	10.647 ug/ml
5) N-HEXADECANE	9.984	1271873	10.582 ug/ml
6) N-OCTADECANE	11.422	1306871	10.691 ug/ml
7) N-EICOSANE	12.729	1297117	10.799 ug/ml
8) N-DOCOSANE	13.925	1298053	10.875 ug/ml
10) N-TETRACOSANE	15.025	1292136	10.968 ug/ml
11) N-HEXACOSANE	16.045	1317703	11.248 ug/ml
12) N-OCTACOSANE	16.992	1265468	11.149 ug/ml
13) N-TRIACONTANE	17.878	1280194	11.405 ug/ml
14) N-DOTRIACONTANE	18.708	1385093	13.044 ug/ml
15) N-TETRATRIACONTANE	19.489	1092522	11.508 ug/ml
16) N-HEXATRIACONTANE	20.226	916153	11.718 ug/ml
17) N-OCTATRIACONTANE	20.932	769712	11.256 ug/ml
18) N-TETRACONTANE	21.770	703096	9.985 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF060622\
Data File : FF010703.D
Signal(s) : FID2B.ch
Acq On : 06 Jun 2022 17:30
Operator : AJ\MA
Sample : 10 TRPH STD
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
10 TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/07/2022
Supervised By :Ankita Jodhani 06/07/2022

Integration File: autoint1.e
Quant Time: Jun 07 03:22:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF060622.M
Quant Title :
QLast Update : Tue Jun 07 03:21:34 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

