

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF060719\
 Data File : FF004752.D
 Signal(s) : FID2B.ch
 Acq On : 07 Jun 2019 20:49
 Operator : DD\AJ
 Sample : K3160-05MS 10X
 Misc :
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :

Manual Integrations
 APPROVED

Ankita
 6/10/2019 11:34:38 AM

Integration File: autoint1.e
 Quant Time: Jun 07 21:43:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
 Quant Title :
 QLast Update : Thu May 23 07:21:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.708	304913	2.497 ug/ml
Target Compounds			
2) N-DECANE	4.416	286572	1.957 ug/ml
3) N-DODECANE	6.439	344173	2.326 ug/ml
4) N-TETRADECANE	8.235	425111	2.894 ug/ml
5) N-HEXADECANE	9.840	561708	3.804 ug/ml
6) N-OCTADECANE	11.286	557226	3.838 ug/ml
7) N-EICOSANE	12.601	540042	3.729 ug/ml
8) N-DOCOSANE	13.805	506587	3.585 ug/ml
10) N-TETRACOSANE	14.911	494606	3.569 ug/ml
11) N-HEXACOSANE	15.939	447850	3.287 ug/ml
12) N-OCTACOSANE	16.894	612965	4.508 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF060719\
Data File : FF004752.D
Signal(s) : FID2B.ch
Acq On : 07 Jun 2019 20:49
Operator : DD\AJ
Sample : K3160-05MS 10X
Misc :
ALS Vial : 67 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :

Manual Integrations
APPROVED

Ankita
6/10/2019 11:34:38 AM

Integration File: autoint1.e
Quant Time: Jun 07 21:43:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
Quant Title :
QLast Update : Thu May 23 07:21:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

