

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF060821\
 Data File : FF008566.D
 Signal(s) : FID2B.ch
 Acq On : 08 Jun 2021 14:38
 Operator : DD\AJ
 Sample : PB136918BS
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 09 01:27:34 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF060621.M
 Quant Title :
 QLast Update : Mon Jun 07 02:02:07 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.768	2026735	19.235 ug/ml
Target Compounds			
1) N-OCTANE	2.419	2029959	17.087 ug/ml
2) N-DECANE	4.527	2349735	19.367 ug/ml
3) N-DODECANE	6.540	2424326	19.432 ug/ml
4) N-TETRADECANE	8.329	2461716	19.344 ug/ml
5) N-HEXADECANE	9.925	2423304	19.171 ug/ml
6) N-OCTADECANE	11.362	2420222	19.094 ug/ml
7) N-EICOSANE	12.668	2434945	19.069 ug/ml
8) N-DOCOSANE	13.867	2399428	19.087 ug/ml
10) N-TETRACOSANE	14.973	2347770	19.160 ug/ml
11) N-HEXACOSANE	15.995	2296815	19.152 ug/ml
12) N-OCTACOSANE	16.943	2263596	19.082 ug/ml
13) N-TRIACONTANE	17.827	2254804	19.038 ug/ml
14) N-DOTRIACONTANE	18.656	2292397	19.049 ug/ml
15) N-TETRATRIACONTANE	19.436	2126593	19.368 ug/ml
16) N-HEXATRIACONTANE	20.171	2056510	18.548 ug/ml
17) N-OCTATRIACONTANE	20.919	1947744	19.194 ug/ml
18) N-TETRACONTANE	21.852	1891579	17.906 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF060821\
Data File : FF008566.D
Signal(s) : FID2B.ch
Acq On : 08 Jun 2021 14:38
Operator : DD\AJ
Sample : PB136918BS
Misc :
ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 09 01:27:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF060621.M
Quant Title :
QLast Update : Mon Jun 07 02:02:07 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

