

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF061518\
 Data File : FF000953.D
 Signal(s) : FID2B.ch
 Acq On : 15 Jun 2018 11:45
 Operator : UA\AJ
 Sample : J3501-01MS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :

Manual Integrations
 APPROVED

ugo
 6/18/2018 11:57:20 AM

Integration File: autoint1.e
 Quant Time: Jun 15 23:41:25 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF060118.M
 Quant Title :
 QLast Update : Fri Jun 01 16:01:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.487	2049683	17.773 ug/ml
Target Compounds			
2) N-DECANE	4.246	2408280	17.188 ug/ml
3) N-DODECANE	6.260	2512346	17.926 ug/ml
4) N-TETRADECANE	8.050	2582475	18.548 ug/ml
5) N-HEXADECANE	9.646	2685357	19.476 ug/ml
6) N-OCTADECANE	11.085	2589454	18.963 ug/ml
7) N-EICOSANE	12.393	2458066	18.347 ug/ml
8) N-DOCOSANE	13.589	2443097	18.476 ug/ml
10) N-TETRACOSANE	14.691	2441325	18.822 ug/ml
11) N-HEXACOSANE	15.710	2356197	18.344 ug/ml
12) N-OCTACOSANE	16.658	2422715	19.053 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF061518\
Data File : FF000953.D
Signal(s) : FID2B.ch
Acq On : 15 Jun 2018 11:45
Operator : UA\AJ
Sample : J3501-01MS
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :

Manual Integrations
APPROVED

ugo
6/18/2018 11:57:20 AM

Integration File: autoint1.e
Quant Time: Jun 15 23:41:25 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF060118.M
Quant Title :
QLast Update : Fri Jun 01 16:01:51 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

