

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF061520\
 Data File : FF007051.D
 Signal(s) : FID2B.ch
 Acq On : 15 Jun 2020 17:53
 Operator : DD\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 16 04:00:45 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF061520.M
 Quant Title :
 QLast Update : Tue Jun 16 03:59:05 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.731	565129	4.826 ug/ml
Target Compounds			
1) N-OCTANE	2.350	640989	4.814 ug/ml
2) N-DECANE	4.461	656539	4.823 ug/ml
3) N-DODECANE	6.483	659254	4.833 ug/ml
4) N-TETRADECANE	8.278	660601	4.874 ug/ml
5) N-HEXADECANE	9.879	657432	4.859 ug/ml
6) N-OCTADECANE	11.322	661814	4.841 ug/ml
7) N-EICOSANE	12.631	658363	4.859 ug/ml
8) N-DOCOSANE	13.830	658620	4.877 ug/ml
10) N-TETRACOSANE	14.934	1061030	7.594 ug/ml
11) N-HEXACOSANE	15.957	768119	5.678 ug/ml
12) N-OCTACOSANE	16.909	856247	6.284 ug/ml
13) N-TRIACONTANE	17.796	874531	6.410 ug/ml
14) N-DOTRIACONTANE	18.630	775026	6.028 ug/ml
15) N-TETRATRIACONTANE	19.413	579471	5.410 ug/ml
16) N-HEXATRIACONTANE	20.155	420279	5.192 ug/ml
17) N-OCTATRIACONTANE	20.907	349579	5.252 ug/ml
18) N-TETRACONTANE	21.843	320837	5.294 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF061520\
Data File : FF007051.D
Signal(s) : FID2B.ch
Acq On : 15 Jun 2020 17:53
Operator : DD\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 62 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 16 04:00:45 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF061520.M
Quant Title :
QLast Update : Tue Jun 16 03:59:05 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

