

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF061819\
Data File : FF004816.D
Signal(s) : FID2B.ch
Acq On : 18 Jun 2019 10:55
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 52 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 19 01:05:46 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
Quant Title :
QLast Update : Thu May 23 07:21:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.704	2049065	16.780 ug/ml
Target Compounds			
16) N-HEXATRIACONTANE	20.194	186566	1.417 ug/ml
18) N-TETRACONTANE	21.875	186286	1.513 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF061819\
Data File : FF004816.D
Signal(s) : FID2B.ch
Acq On : 18 Jun 2019 10:55
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 52 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 19 01:05:46 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
Quant Title :
QLast Update : Thu May 23 07:21:19 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.704 min
Delta R.T.: -0.004 min
Response: 2049065
Conc: 16.78 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.194 min
Delta R.T.: 0.044 min
Response: 186566
Conc: 1.42 ug/ml



#18 N-TETRACONTANE

R.T.: 21.875 min
Delta R.T.: 0.038 min
Response: 186286
Conc: 1.51 ug/ml