

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF062222\
 Data File : FF010813.D
 Signal(s) : FID2B.ch
 Acq On : 22 Jun 2022 15:32
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Jun 23 01:09:45 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF060622.M
 Quant Title :
 QLast Update : Tue Jun 07 09:12:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.823	5669694	50.158 ug/ml
Target Compounds			
2) N-DECANE	4.384	6244130	52.232 ug/ml
3) N-DODECANE	6.551	6276291	51.645 ug/ml
4) N-TETRADECANE	8.375	6380475	51.465 ug/ml
5) N-HEXADECANE	9.979	6454592	51.621 ug/ml
6) N-OCTADECANE	11.419	6561319	51.276 ug/ml
7) N-EICOSANE	12.727	6466113	51.130 ug/ml
8) N-DOCOSANE	13.924	6436490	50.964 ug/ml
10) N-TETRACOSANE	15.027	6313017	50.365 ug/ml
11) N-HEXACOSANE	16.046	6294469	49.693 ug/ml
12) N-OCTACOSANE	16.994	6169741	50.556 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF062222\
Data File : FF010813.D
Signal(s) : FID2B.ch
Acq On : 22 Jun 2022 15:32
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Jun 23 01:09:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF060622.M
Quant Title :
QLast Update : Tue Jun 07 09:12:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

