

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF062222\
 Data File : FF010819.D
 Signal(s) : FID2B.ch
 Acq On : 22 Jun 2022 18:27
 Operator : AJ\MA
 Sample : N3422-05MSD
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 WC-14A-1A-12-GRMSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/23/2022
 Supervised By :Ankita Jodhani 06/23/2022

Integration File: autoint1.e
 Quant Time: Jun 23 01:50:45 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF060622.M
 Quant Title :
 QLast Update : Tue Jun 07 09:12:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.818	1304445	11.540 ug/mlm
Target Compounds			
2) N-DECANE	4.388	3443501	28.805 ug/ml
3) N-DODECANE	6.555	3040156	25.016 ug/ml
4) N-TETRADECANE	8.381	2153439	17.370 ug/mlm
5) N-HEXADECANE	9.989	3573394	28.578 ug/mlm
6) N-OCTADECANE	11.439	3524942	27.547 ug/mlm
7) N-EICOSANE	12.732	326958	2.585 ug/mlm
8) N-DOCOSANE	13.925	76291	0.604 ug/mlm
10) N-TETRACOSANE	15.006	1327173	10.588 ug/mlm
11) N-HEXACOSANE	16.040	48963	0.387 ug/mlm
12) N-OCTACOSANE	16.996	116964	0.958 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF062222\
Data File : FF010819.D
Signal(s) : FID2B.ch
Acq On : 22 Jun 2022 18:27
Operator : AJ\MA
Sample : N3422-05MSD
Misc :
ALS Vial : 59 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
WC-14A-1A-12-GRMSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/23/2022
Supervised By :Ankita Jodhani 06/23/2022

Integration File: autoint1.e
Quant Time: Jun 23 01:50:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF060622.M
Quant Title :
QLast Update : Tue Jun 07 09:12:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

