

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_F\Data\FF062619\  
 Data File : FF004851.D  
 Signal(s) : FID2B.ch  
 Acq On : 26 Jun 2019 11:40  
 Operator : DD\AJ  
 Sample : I.BLK  
 Misc :  
 ALS Vial : 52 Sample Multiplier: 1

Instrument :  
 FID\_F  
 ClientSampleId :

Integration File: autoint1.e  
 Quant Time: Jun 26 13:24:32 2019  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_F\Method\FF062619.M  
 Quant Title :  
 QLast Update : Wed Jun 26 13:23:59 2019  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1uL  
 Signal Phase : Rxi-1ms  
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units
-----			
System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.691	1952702	18.982 ug/ml
Target Compounds			
16) N-HEXATRIACONTANE	20.188	148335	1.373 ug/ml
18) N-TETRACONTANE	21.857	239293	2.203 ug/ml
-----			

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_F\Data\FF062619\  
Data File : FF004851.D  
Signal(s) : FID2B.ch  
Acq On : 26 Jun 2019 11:40  
Operator : DD\AJ  
Sample : I.BLK  
Misc :  
ALS Vial : 52 Sample Multiplier: 1

Instrument :  
FID\_F  
ClientSampled :

Integration File: autoint1.e  
Quant Time: Jun 26 13:24:32 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_F\Method\FF062619.M  
Quant Title :  
QLast Update : Wed Jun 26 13:23:59 2019  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1uL  
Signal Phase : Rxi-1ms  
Signal Info : 20mx0.18mmx0.18um






#9 TETRACOSANE-d50 (SURROGATE)


R.T.: 14.691 min  
 Delta R.T.: 0.000 min  
 Response: 1952702  
 Conc: 18.98 ug/ml

Instrument :  
 FID\_F  
 ClientSampleId :



#16 N-HEXATRIACONTANE

R.T.: 20.188 min  
 Delta R.T.: 0.046 min  
 Response: 148335  
 Conc: 1.37 ug/ml



#18 N-TETRACONTANE

R.T.: 21.857 min  
 Delta R.T.: 0.039 min  
 Response: 239293  
 Conc: 2.20 ug/ml