

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF062818\
 Data File : FF001142.D
 Signal(s) : FID2B.ch
 Acq On : 28 Jun 2018 19:01
 Operator : UA\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jun 29 01:23:27 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF060118.M
 Quant Title :
 QLast Update : Fri Jun 01 16:01:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.471	5380514	46.656 ug/ml
Target Compounds			
2) N-DECANE	4.236	8137348	58.075 ug/ml
3) N-DODECANE	6.249	7670001	54.726 ug/ml
4) N-TETRADECANE	8.037	7233581	51.954 ug/ml
5) N-HEXADECANE	9.633	6868075	49.813 ug/ml
6) N-OCTADECANE	11.071	6592013	48.274 ug/ml
7) N-EICOSANE	12.378	6322578	47.191 ug/ml
8) N-DOCOSANE	13.573	6155972	46.556 ug/ml
10) N-TETRACOSANE	14.675	6076349	46.846 ug/ml
11) N-HEXACOSANE	15.694	6039864	47.023 ug/ml
12) N-OCTACOSANE	16.642	6128814	48.198 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF062818\
Data File : FF001142.D
Signal(s) : FID2B.ch
Acq On : 28 Jun 2018 19:01
Operator : UA\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jun 29 01:23:27 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF060118.M
Quant Title :
QLast Update : Fri Jun 01 16:01:51 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

