

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF063022\
 Data File : FF010932.D
 Signal(s) : FID2B.ch
 Acq On : 30 Jun 2022 12:37
 Operator : AJ\MA
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 30 13:00:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
 Quant Title :
 QLast Update : Thu Jun 30 13:00:03 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.817	521105	5.713 ug/ml
Target Compounds			
1) N-OCTANE	1.868	628636	5.734 ug/ml
2) N-DECANE	4.376	618632	5.622 ug/ml
3) N-DODECANE	6.543	605168	5.554 ug/ml
4) N-TETRADECANE	8.368	608793	5.568 ug/ml
5) N-HEXADECANE	9.972	599519	5.557 ug/ml
6) N-OCTADECANE	11.413	602844	5.593 ug/ml
7) N-EICOSANE	12.721	589084	5.614 ug/mlm
8) N-DOCOSANE	13.917	584668	5.667 ug/ml
10) N-TETRACOSANE	15.020	580882	5.695 ug/ml
11) N-HEXACOSANE	16.040	581688	5.750 ug/ml
12) N-OCTACOSANE	16.989	588273	5.822 ug/ml
13) N-TRIACONTANE	17.874	596437	5.882 ug/mlm
14) N-DOTRIACONTANE	18.705	632935	6.174 ug/ml
15) N-TETRATRIACONTANE	19.487	557717	5.934 ug/mlm
16) N-HEXATRIACONTANE	20.223	485546	5.836 ug/mlm
17) N-OCTATRIACONTANE	20.931	433798	6.022 ug/mlm
18) N-TETRACONTANE	21.761	479833	5.815 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF063022\
Data File : FF010932.D
Signal(s) : FID2B.ch
Acq On : 30 Jun 2022 12:37
Operator : AJ\MA
Sample : 5 TRPH STD
Misc :
ALS Vial : 57 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 30 13:00:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
Quant Title :
QLast Update : Thu Jun 30 13:00:03 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

