

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF063022\
 Data File : FF010933.D
 Signal(s) : FID2B.ch
 Acq On : 30 Jun 2022 13:49
 Operator : AJ\MA
 Sample : FF063022ICV
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jun 30 14:14:41 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
 Quant Title :
 QLast Update : Thu Jun 30 13:41:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.820	4382953	46.722 ug/ml
Target Compounds			
1) N-OCTANE	1.859	5272776	46.726 ug/ml
2) N-DECANE	4.371	5305792	47.048 ug/ml
3) N-DODECANE	6.542	5277275	47.385 ug/ml
4) N-TETRADECANE	8.368	5282069	47.236 ug/ml
5) N-HEXADECANE	9.974	5220503	47.331 ug/ml
6) N-OCTADECANE	11.415	5223287	47.339 ug/ml
7) N-EICOSANE	12.724	5060257	46.980 ug/ml
8) N-DOCOSANE	13.921	4974112	46.961 ug/ml
10) N-TETRACOSANE	15.023	4901395	46.752 ug/ml
11) N-HEXACOSANE	16.043	4870065	46.737 ug/ml
12) N-OCTACOSANE	16.992	4845232	46.425 ug/ml
13) N-TRIACONTANE	17.878	4845497	46.099 ug/ml
14) N-DOTRIACONTANE	18.709	4874812	45.421 ug/ml
15) N-TETRATRIACONTANE	19.491	4498823	46.032 ug/ml
16) N-HEXATRIACONTANE	20.229	4001126	47.762 ug/ml
17) N-OCTATRIACONTANE	20.936	3516043	47.809 ug/ml
18) N-TETRACONTANE	21.765	3780435	44.368 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF063022\
Data File : FF010933.D
Signal(s) : FID2B.ch
Acq On : 30 Jun 2022 13:49
Operator : AJ\MA
Sample : FF063022ICV
Misc :
ALS Vial : 59 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jun 30 14:14:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
Quant Title :
QLast Update : Thu Jun 30 13:41:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

