

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070119\
 Data File : FF004902.D
 Signal(s) : FID2B.ch
 Acq On : 02 Jul 2019 00:32
 Operator : DD\AJ
 Sample : K3569-06
 Misc :
 ALS Vial : 72 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 LT-3

Integration File: autoint1.e
 Quant Time: Jul 02 01:54:08 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062619.M
 Quant Title :
 QLast Update : Wed Jun 26 14:27:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.690	2097838	21.570 ug/ml
Target Compounds			
13) N-TRIACONTANE	17.768	95775	0.853 ug/ml
14) N-DOTRIACONTANE	18.609	153948	1.352 ug/ml
16) N-HEXATRIACONTANE	20.180	163006	1.574 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070119\
Data File : FF004902.D
Signal(s) : FID2B.ch
Acq On : 02 Jul 2019 00:32
Operator : DD\AJ
Sample : K3569-06
Misc :
ALS Vial : 72 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
LT-3

Integration File: autoint1.e
Quant Time: Jul 02 01:54:08 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062619.M
Quant Title :
QLast Update : Wed Jun 26 14:27:24 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#9 TETRACOSANE-d50 (SURROGATE)


R.T.: 14.690 min
Delta R.T.: -0.002 min
Response: 2097838
Conc: 21.57 ug/ml

Instrument :
FID_F
ClientSampleId :
LT-3




#13 N-TRIACONTANE

R.T.: 17.768 min
Delta R.T.: -0.002 min
Response: 95775
Conc: 0.85 ug/ml



#14 N-DOTRIACONTANE

R.T.: 18.609 min
Delta R.T.: 0.002 min
Response: 153948
Conc: 1.35 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.180 min
Delta R.T.: 0.039 min
Response: 163006
Conc: 1.57 ug/ml