

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070119\
 Data File : FF004892.D
 Signal(s) : FID2B.ch
 Acq On : 01 Jul 2019 19:37
 Operator : DD\AJ
 Sample : PB121008BS
 Misc :
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 PB121008BS

Integration File: autoint1.e
 Quant Time: Jul 02 01:51:40 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062619.M
 Quant Title :
 QLast Update : Wed Jun 26 14:27:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.694	2628329	27.024 ug/ml
Target Compounds			
2) N-DECANE	4.405	3256208	25.785 ug/ml
3) N-DODECANE	6.428	3299411	26.004 ug/ml
4) N-TETRADECANE	8.223	3285859	26.260 ug/ml
5) N-HEXADECANE	9.826	3248561	26.309 ug/ml
6) N-OCTADECANE	11.272	3188476	26.914 ug/ml
7) N-EICOSANE	12.586	3132958	27.002 ug/ml
8) N-DOCOSANE	13.789	3087872	27.444 ug/ml
10) N-TETRACOSANE	14.899	3119630	28.025 ug/ml
11) N-HEXACOSANE	15.925	3118957	28.199 ug/ml
12) N-OCTACOSANE	16.880	3175779	28.472 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070119\
Data File : FF004892.D
Signal(s) : FID2B.ch
Acq On : 01 Jul 2019 19:37
Operator : DD\AJ
Sample : PB121008BS
Misc :
ALS Vial : 64 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
PB121008BS

Integration File: autoint1.e
Quant Time: Jul 02 01:51:40 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062619.M
Quant Title :
QLast Update : Wed Jun 26 14:27:24 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

