

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070122\
 Data File : FF010952.D
 Signal(s) : FID2B.ch
 Acq On : 01 Jul 2022 13:03
 Operator : AJ\MA
 Sample : N3214-06
 Misc : 1 PPM SOIL MDL
 ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 02 02:57:49 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
 Quant Title :
 QLast Update : Fri Jul 01 01:40:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.816	2533732	27.009 ug/ml
Target Compounds			
1) N-OCTANE	1.862	129585	1.148 ug/ml
2) N-DECANE	4.368	113375	1.005 ug/ml
3) N-DODECANE	6.539	103722	0.931 ug/ml
4) N-TETRADECANE	8.366	112318	1.004 ug/ml
5) N-HEXADECANE	9.973	103993	0.943 ug/ml
6) N-OCTADECANE	11.411	108781	0.986 ug/ml
7) N-EICOSANE	12.719	117169	1.090 ug/ml
8) N-DOCOSANE	13.914	119510	1.128 ug/mlm
10) N-TETRACOSANE	15.018	123105	1.174 ug/ml
11) N-HEXACOSANE	16.038	117621	1.129 ug/ml
12) N-OCTACOSANE	16.987	117154	1.123 ug/ml
13) N-TRIACONTANE	17.873	121682	1.159 ug/ml
14) N-DOTRIACONTANE	18.704	140925	1.313 ug/ml
15) N-TETRATRIACONTANE	19.486	87880	0.901 ug/mlm
16) N-HEXATRIACONTANE	20.224	75608	0.879 ug/mlm
17) N-OCTATRIACONTANE	20.931	66403	0.886 ug/mlm
18) N-TETRACONTANE	21.761	132215	1.552 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070122\
Data File : FF010952.D
Signal(s) : FID2B.ch
Acq On : 01 Jul 2022 13:03
Operator : AJ\MA
Sample : N3214-06
Misc : 1 PPM SOIL MDL
ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 02 02:57:49 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
Quant Title :
QLast Update : Fri Jul 01 01:40:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

