

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070122\
 Data File : FF010955.D
 Signal(s) : FID2B.ch
 Acq On : 01 Jul 2022 14:37
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 02 02:58:23 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
 Quant Title :
 QLast Update : Fri Jul 01 01:40:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.817	4874814	51.965 ug/ml
Target Compounds			
1) N-OCTANE	1.866	5126842	45.433 ug/ml
2) N-DECANE	4.376	5238086	46.447 ug/ml
3) N-DODECANE	6.545	5293582	47.532 ug/ml
4) N-TETRADECANE	8.369	5437071	48.622 ug/ml
5) N-HEXADECANE	9.974	5506397	49.924 ug/ml
6) N-OCTADECANE	11.414	5623982	50.970 ug/ml
7) N-EICOSANE	12.722	5534157	51.480 ug/ml
8) N-DOCOSANE	13.918	5495654	51.885 ug/ml
10) N-TETRACOSANE	15.020	5434395	51.836 ug/ml
11) N-HEXACOSANE	16.040	5349817	51.341 ug/ml
12) N-OCTACOSANE	16.989	5235457	50.164 ug/ml
13) N-TRIACONTANE	17.874	5135345	48.920 ug/ml
14) N-DOTRIACONTANE	18.706	4860722	45.289 ug/ml
15) N-TETRATRIACONTANE	19.487	4312590	44.231 ug/mlm
16) N-HEXATRIACONTANE	20.223	3547607	41.259 ug/mlm
17) N-OCTATRIACONTANE	20.929	2959060	39.465 ug/ml
18) N-TETRACONTANE	21.755	2955353	34.684 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070122\
Data File : FF010955.D
Signal(s) : FID2B.ch
Acq On : 01 Jul 2022 14:37
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 02 02:58:23 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
Quant Title :
QLast Update : Fri Jul 01 01:40:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

