

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070124\
 Data File : FF014296.D
 Signal(s) : FID2B.ch
 Acq On : 01 Jul 2024 12:59
 Operator : YP\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 65 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 5 TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/01/2024
 Supervised By :Ankita Jodhani 07/02/2024

Integration File: autoint1.e
 Quant Time: Jul 01 13:19:38 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF070124.M
 Quant Title :
 QLast Update : Mon Jul 01 13:19:20 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.925	619273	5.418 ug/ml
Target Compounds			
1) N-OCTANE	1.966	684658	5.518 ug/ml
2) N-DECANE	4.476	658849	5.374 ug/ml
3) N-DODECANE	6.641	652221	5.252 ug/ml
4) N-TETRADECANE	8.467	619563	5.131 ug/ml
5) N-HEXADECANE	10.073	617586	5.044 ug/ml
6) N-OCTADECANE	11.515	673719	5.242 ug/ml
7) N-EICOSANE	12.826	670934	5.302 ug/ml
8) N-DOCOSANE	14.023	667590	5.362 ug/ml
10) N-TETRACOSANE	15.128	665308	5.407 ug/ml
11) N-HEXACOSANE	16.150	636255	5.361 ug/ml
12) N-OCTACOSANE	17.101	622738	5.403 ug/ml
13) N-TRIACONTANE	17.988	606086	5.422 ug/ml
14) N-DOTRIACONTANE	18.822	591786	5.565 ug/ml
15) N-TETRATRIACONTANE	19.604	469849	5.459 ug/ml
16) N-HEXATRIACONTANE	20.344	345399	5.174 ug/mlm
17) N-OCTATRIACONTANE	21.068	323463	5.278 ug/ml
18) N-TETRACONTANE	21.942	333767	5.764 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070124\
Data File : FF014296.D
Signal(s) : FID2B.ch
Acq On : 01 Jul 2024 12:59
Operator : YP\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 65 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
5 TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/01/2024
Supervised By :Ankita Jodhani 07/02/2024

Integration File: autoint1.e
Quant Time: Jul 01 13:19:38 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF070124.M
Quant Title :
QLast Update : Mon Jul 01 13:19:20 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

