

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070218\
 Data File : FF001188.D
 Signal(s) : FID2B.ch
 Acq On : 02 Jul 2018 13:13
 Operator : UA\AJ
 Sample : 10 TRPH STD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 02 13:42:28 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF070218.M
 Quant Title :
 QLast Update : Mon Jul 02 13:22:41 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.463	1079627	10.030 ug/ml
Target Compounds			
1) N-OCTANE	2.150	1209600	9.626 ug/ml
2) N-DECANE	4.229	1202961	9.590 ug/ml
3) N-DODECANE	6.241	1224899	9.703 ug/ml
4) N-TETRADECANE	8.030	1241516	9.829 ug/ml
5) N-HEXADECANE	9.626	1226074	9.752 ug/ml
6) N-OCTADECANE	11.063	1239561	9.855 ug/ml
7) N-EICOSANE	12.370	1242258	9.982 ug/ml
8) N-DOCOSANE	13.565	1237068	10.043 ug/ml
10) N-TETRACOSANE	14.666	1222040	10.032 ug/ml
11) N-HEXACOSANE	15.684	1215420	10.036 ug/ml
12) N-OCTACOSANE	16.633	1208077	10.037 ug/ml
13) N-TRIACONTANE	17.517	1210898	10.098 ug/ml
14) N-DOTRIACONTANE	18.350	1188695	10.110 ug/ml
15) N-TETRATRIACONTANE	19.132	1172865	9.972 ug/ml
16) N-HEXATRIACONTANE	19.873	1125987	9.884 ug/ml
17) N-OCTATRIACONTANE	20.585	1022578	9.872 ug/ml
18) N-TETRACONTANE	21.417	987345	9.528 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070218\
Data File : FF001188.D
Signal(s) : FID2B.ch
Acq On : 02 Jul 2018 13:13
Operator : UA\AJ
Sample : 10 TRPH STD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 02 13:42:28 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF070218.M
Quant Title :
QLast Update : Mon Jul 02 13:22:41 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

