

Method Path : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\
 Method File : FF070218.M
 Title :
 Last Update : Mon Jul 02 13:42:46 2018
 Response Via : Initial Calibration

Calibration Files

50 =FF001186.D 100 =FF001185.D 20 =FF001187.D
 10 =FF001188.D 5 =FF000702.D

	Compound	50	100	20	10	5	Avg	%RSD
1)	N-OCTANE	1.279	1.176	1.315	1.210		1.245 E5	5.11
2)	N-DECANE	1.277	1.176	1.310	1.203		1.242 E5	5.05
3)	N-DODECANE	1.282	1.179	1.326	1.225		1.253 E5	5.13
4)	N-TETRADECANE	1.281	1.176	1.332	1.242		1.258 E5	5.23
5)	N-HEXADECANE	1.276	1.174	1.322	1.226		1.249 E5	5.12
6)	N-OCTADECANE	1.276	1.173	1.325	1.240		1.253 E5	5.11
7)	N-EICOSANE	1.262	1.160	1.312	1.242		1.244 E5	5.08
8)	N-DOCOSANE	1.249	1.149	1.298	1.237		1.233 E5	5.03
9) S	TETRACOSANE-d50 (S...	1.089	1.002	1.138	1.080		1.077 E5	5.22
10)	N-TETRACOSANE	1.235	1.137	1.283	1.222		1.219 E5	4.99
11)	N-HEXACOSANE	1.227	1.131	1.275	1.215		1.212 E5	4.93
12)	N-OCTACOSANE	1.217	1.126	1.268	1.208		1.205 E5	4.90
13)	N-TRIACONTANE	1.208	1.122	1.267	1.211		1.202 E5	4.95
14)	N-DOTRIACONTANE	1.181	1.101	1.246	1.189		1.179 E5	5.06
15)	N-TETRATRIACONTANE	1.178	1.107	1.243	1.173		1.175 E5	4.72
16)	N-HEXATRIACONTANE	1.140	1.078	1.200	1.126		1.136 E5	4.42
17)	N-OCTATRIACONTANE	1.033	0.988	1.087	1.023		1.033 E5	3.99
18)	N-TETRACONTANE	1.031	0.990	1.087	0.987		1.024 E5	4.55

(#) = Out of Range