

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070622\
 Data File : FF011010.D
 Signal(s) : FID2B.ch
 Acq On : 06 Jul 2022 15:46
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 07 01:09:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
 Quant Title :
 QLast Update : Fri Jul 01 01:40:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.812	5153035	54.931 ug/ml
Target Compounds			
2) N-DECANE	4.371	5840987	51.793 ug/ml
3) N-DODECANE	6.540	5907453	53.044 ug/ml
4) N-TETRADECANE	8.364	6019254	53.828 ug/ml
5) N-HEXADECANE	9.969	6049943	54.852 ug/ml
6) N-OCTADECANE	11.409	6108932	55.365 ug/ml
7) N-EICOSANE	12.717	5954546	55.391 ug/ml
8) N-DOCOSANE	13.914	5855801	55.285 ug/ml
10) N-TETRACOSANE	15.015	5705802	54.425 ug/ml
11) N-HEXACOSANE	16.035	5551479	53.276 ug/ml
12) N-OCTACOSANE	16.984	5367497	51.429 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070622\
Data File : FF011010.D
Signal(s) : FID2B.ch
Acq On : 06 Jul 2022 15:46
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 07 01:09:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
Quant Title :
QLast Update : Fri Jul 01 01:40:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

