

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070622\
 Data File : FF011013.D
 Signal(s) : FID2B.ch
 Acq On : 06 Jul 2022 17:38
 Operator : AJ\MA
 Sample : N3214-07
 Misc : 2.5 PPM WATER LOD
 ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 07 01:09:36 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
 Quant Title :
 QLast Update : Fri Jul 01 01:40:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.810	2170168	23.134 ug/ml
Target Compounds			
2) N-DECANE	4.371	312087	2.767 ug/ml
3) N-DODECANE	6.539	307733	2.763 ug/ml
4) N-TETRADECANE	8.364	322041	2.880 ug/ml
5) N-HEXADECANE	9.968	316131	2.866 ug/ml
6) N-OCTADECANE	11.407	325324	2.948 ug/ml
7) N-EICOSANE	12.714	321846	2.994 ug/ml
8) N-DOCOSANE	13.910	314935	2.973 ug/ml
10) N-TETRACOSANE	15.012	314080	2.996 ug/ml
11) N-HEXACOSANE	16.033	299670	2.876 ug/ml
12) N-OCTACOSANE	16.981	293915	2.816 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070622\
Data File : FF011013.D
Signal(s) : FID2B.ch
Acq On : 06 Jul 2022 17:38
Operator : AJ\MA
Sample : N3214-07
Misc : 2.5 PPM WATER LOD
ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 07 01:09:36 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
Quant Title :
QLast Update : Fri Jul 01 01:40:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

