

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070622\
 Data File : FF011016.D
 Signal(s) : FID2B.ch
 Acq On : 06 Jul 2022 19:08
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 07 01:10:10 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
 Quant Title :
 QLast Update : Fri Jul 01 01:40:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.812	5135288	54.742 ug/ml
Target Compounds			
2) N-DECANE	4.372	5808744	51.507 ug/ml
3) N-DODECANE	6.541	5865663	52.669 ug/ml
4) N-TETRADECANE	8.366	5980574	53.483 ug/ml
5) N-HEXADECANE	9.970	6009483	54.485 ug/ml
6) N-OCTADECANE	11.410	6066699	54.983 ug/ml
7) N-EICOSANE	12.718	5920337	55.072 ug/ml
8) N-DOCOSANE	13.914	5826038	55.004 ug/ml
10) N-TETRACOSANE	15.016	5690016	54.274 ug/ml
11) N-HEXACOSANE	16.036	5548935	53.252 ug/ml
12) N-OCTACOSANE	16.984	5371891	51.471 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070622\
Data File : FF011016.D
Signal(s) : FID2B.ch
Acq On : 06 Jul 2022 19:08
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 07 01:10:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
Quant Title :
QLast Update : Fri Jul 01 01:40:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

