

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070722\
 Data File : FF011028.D
 Signal(s) : FID2B.ch
 Acq On : 07 Jul 2022 16:20
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 08 03:15:44 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
 Quant Title :
 QLast Update : Fri Jul 01 01:40:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.811	5431709	57.901 ug/ml
Target Compounds			
2) N-DECANE	4.371	6121273	54.279 ug/ml
3) N-DODECANE	6.540	6192751	55.606 ug/ml
4) N-TETRADECANE	8.364	6313805	56.463 ug/ml
5) N-HEXADECANE	9.968	6356191	57.628 ug/ml
6) N-OCTADECANE	11.408	6419736	58.182 ug/ml
7) N-EICOSANE	12.716	6263558	58.265 ug/ml
8) N-DOCOSANE	13.912	6160229	58.159 ug/ml
10) N-TETRACOSANE	15.014	6013714	57.362 ug/ml
11) N-HEXACOSANE	16.034	5836469	56.011 ug/ml
12) N-OCTACOSANE	16.981	5627590	53.921 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF070722\
Data File : FF011028.D
Signal(s) : FID2B.ch
Acq On : 07 Jul 2022 16:20
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 08 03:15:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
Quant Title :
QLast Update : Fri Jul 01 01:40:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

