

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF071019\
 Data File : FF004999.D
 Signal(s) : FID2B.ch
 Acq On : 10 Jul 2019 10:57
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Jul 11 00:45:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062619.M
 Quant Title :
 QLast Update : Wed Jun 26 14:27:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.681	4973537	51.137 ug/ml
Target Compounds			
2) N-DECANE	4.397	6165643	48.825 ug/ml
3) N-DODECANE	6.420	6192780	48.807 ug/ml
4) N-TETRADECANE	8.214	6101554	48.763 ug/ml
5) N-HEXADECANE	9.817	6018211	48.739 ug/ml
6) N-OCTADECANE	11.261	5811659	49.056 ug/ml
7) N-EICOSANE	12.576	5745894	49.523 ug/ml
8) N-DOCOSANE	13.778	5662557	50.327 ug/ml
10) N-TETRACOSANE	14.886	5690195	51.118 ug/ml
11) N-HEXACOSANE	15.912	5696098	51.499 ug/ml
12) N-OCTACOSANE	16.867	5726560	51.341 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF071019\
Data File : FF004999.D
Signal(s) : FID2B.ch
Acq On : 10 Jul 2019 10:57
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Jul 11 00:45:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062619.M
Quant Title :
QLast Update : Wed Jun 26 14:27:24 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

