

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF071119\
Data File : FF005020.D
Signal(s) : FID2B.ch
Acq On : 11 Jul 2019 09:20
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 11 12:20:46 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062619.M
Quant Title :
QLast Update : Wed Jun 26 14:27:24 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.691	5033113	51.750 ug/ml
Target Compounds			
2) N-DECANE	4.399	5919319	46.874 ug/ml
3) N-DODECANE	6.423	5986901	47.185 ug/ml
4) N-TETRADECANE	8.219	5952574	47.572 ug/ml
5) N-HEXADECANE	9.823	5943991	48.138 ug/ml
6) N-OCTADECANE	11.268	5802027	48.975 ug/ml
7) N-EICOSANE	12.583	5777419	49.795 ug/ml
8) N-DOCOSANE	13.787	5714678	50.790 ug/ml
10) N-TETRACOSANE	14.895	5758033	51.727 ug/ml
11) N-HEXACOSANE	15.922	5774977	52.212 ug/ml
12) N-OCTACOSANE	16.878	5843061	52.386 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF071119\
Data File : FF005020.D
Signal(s) : FID2B.ch
Acq On : 11 Jul 2019 09:20
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 11 12:20:46 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062619.M
Quant Title :
QLast Update : Wed Jun 26 14:27:24 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

