

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF071119\
 Data File : FF005028.D
 Signal(s) : FID2B.ch
 Acq On : 11 Jul 2019 13:50
 Operator : DD\AJ
 Sample : K3591-07
 Misc : LOD-WATER- 2.5 PPM
 ALS Vial : 61 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 11 15:34:34 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062619.M
 Quant Title :
 QLast Update : Wed Jun 26 14:27:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.683	2223937	22.866 ug/ml
Target Compounds			
2) N-DECANE	4.395	321344	2.545 ug/ml
3) N-DODECANE	6.418	323740	2.551 ug/ml
4) N-TETRADECANE	8.213	310754	2.483 ug/ml
5) N-HEXADECANE	9.815	326633	2.645 ug/ml
6) N-OCTADECANE	11.261	325591	2.748 ug/ml
7) N-EICOSANE	12.575	324866	2.800 ug/ml
8) N-DOCOSANE	13.778	321999	2.862 ug/ml
10) N-TETRACOSANE	14.886	336683	3.025 ug/ml
11) N-HEXACOSANE	15.913	329895	2.983 ug/ml
12) N-OCTACOSANE	16.868	355342	3.186 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF071119\
Data File : FF005028.D
Signal(s) : FID2B.ch
Acq On : 11 Jul 2019 13:50
Operator : DD\AJ
Sample : K3591-07
Misc : LOD-WATER- 2.5 PPM
ALS Vial : 61 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 11 15:34:34 2019
Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062619.M
Quant Title :
QLast Update : Wed Jun 26 14:27:24 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

