

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF071119\
 Data File : FF005029.D
 Signal(s) : FID2B.ch
 Acq On : 11 Jul 2019 14:20
 Operator : DD\AJ
 Sample : K3591-08
 Misc : LOQ-WATER- 5 PPM
 ALS Vial : 62 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 11 15:34:40 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062619.M
 Quant Title :
 QLast Update : Wed Jun 26 14:27:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.683	2019239	20.762 ug/ml
Target Compounds			
2) N-DECANE	4.395	590670	4.677 ug/ml
3) N-DODECANE	6.418	598432	4.716 ug/ml
4) N-TETRADECANE	8.213	585127	4.676 ug/ml
5) N-HEXADECANE	9.816	597286	4.837 ug/ml
6) N-OCTADECANE	11.261	589324	4.974 ug/ml
7) N-EICOSANE	12.576	587215	5.061 ug/ml
8) N-DOCOSANE	13.779	580094	5.156 ug/ml
10) N-TETRACOSANE	14.887	599169	5.383 ug/ml
11) N-HEXACOSANE	15.913	586019	5.298 ug/ml
12) N-OCTACOSANE	16.868	603638	5.412 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF071119\
Data File : FF005029.D
Signal(s) : FID2B.ch
Acq On : 11 Jul 2019 14:20
Operator : DD\AJ
Sample : K3591-08
Misc : LOQ-WATER- 5 PPM
ALS Vial : 62 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 11 15:34:40 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062619.M
Quant Title :
QLast Update : Wed Jun 26 14:27:24 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

