

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF071221\
 Data File : FF008775.D
 Signal(s) : FID2B.ch
 Acq On : 12 Jul 2021 15:57
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Jul 13 01:26:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062321.M
 Quant Title :
 QLast Update : Thu Jun 24 08:07:11 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.752	4610434	40.453 ug/ml
Target Compounds				
2)	N-DECANE	4.506	6396690	47.444 ug/ml
3)	N-DODECANE	6.524	6296631	46.297 ug/ml
4)	N-TETRADECANE	8.316	6147935	45.568 ug/ml
5)	N-HEXADECANE	9.913	5932701	43.428 ug/ml
6)	N-OCTADECANE	11.352	5874403	42.209 ug/ml
7)	N-EICOSANE	12.660	5664696	41.314 ug/ml
8)	N-DOCOSANE	13.854	5444032	40.688 ug/ml
10)	N-TETRACOSANE	14.957	5334763	40.486 ug/ml
11)	N-HEXACOSANE	15.977	5191646	40.630 ug/ml
12)	N-OCTACOSANE	16.926	5139932	41.556 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF071221\
Data File : FF008775.D
Signal(s) : FID2B.ch
Acq On : 12 Jul 2021 15:57
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Jul 13 01:26:56 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062321.M
Quant Title :
QLast Update : Thu Jun 24 08:07:11 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

