

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF071818\
 Data File : FF001404.D
 Signal(s) : FID2B.ch
 Acq On : 18 Jul 2018 13:33
 Operator : AJ\SJ
 Sample : J4035-04
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 19 02:09:12 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF070218.M
 Quant Title :
 QLast Update : Mon Jul 02 13:42:46 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.455	2309224	21.437 ug/ml
Target Compounds			
4) N-TETRADECANE	8.076	465265	3.699 ug/ml
11) N-HEXACOSANE	15.684	83137	0.686 ug/ml
12) N-OCTACOSANE	16.636	475470	3.947 ug/ml
13) N-TRIACONTANE	17.504	810015	6.738 ug/ml
14) N-DOTRIACONTANE	18.350	218579	1.854 ug/ml
15) N-TETRATRIACONTANE	19.148	562626	4.787 ug/ml
16) N-HEXATRIACONTANE	19.874	277692	2.445 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF071818\
Data File : FF001404.D
Signal(s) : FID2B.ch
Acq On : 18 Jul 2018 13:33
Operator : AJ\SJ
Sample : J4035-04
Misc :
ALS Vial : 55 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 19 02:09:12 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF070218.M
Quant Title :
QLast Update : Mon Jul 02 13:42:46 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#4 N-TETRADECANE

R.T.: 8.076 min
Delta R.T.: 0.043 min
Response: 465265
Conc: 3.70 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.455 min
Delta R.T.: -0.012 min
Response: 2309224
Conc: 21.44 ug/ml




#11 N-HEXACOSANE

R.T.: 15.684 min
Delta R.T.: -0.005 min
Response: 83137
Conc: 0.69 ug/ml




#12 N-OCTACOSANE

R.T.: 16.636 min
Delta R.T.: -0.002 min
Response: 475470
Conc: 3.95 ug/ml




#13 N-TRIACONTANE

R.T.: 17.504 min
Delta R.T.: -0.019 min
Response: 810015
Conc: 6.74 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.350 min
Delta R.T.: -0.003 min
Response: 218579
Conc: 1.85 ug/ml



#15 N-TETRATRIACONTANE

R.T.: 19.148 min
Delta R.T.: 0.011 min
Response: 562626
Conc: 4.79 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 19.874 min
Delta R.T.: -0.003 min
Response: 277692
Conc: 2.44 ug/ml