

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF072021\
 Data File : FF008860.D
 Signal(s) : FID2B.ch
 Acq On : 20 Jul 2021 14:02
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 21 01:09:09 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062321.M
 Quant Title :
 QLast Update : Thu Jun 24 08:07:11 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.744	4766822	41.826 ug/ml
Target Compounds			
1) N-OCTANE	2.398	6808961	50.815 ug/ml
2) N-DECANE	4.505	6861147	50.888 ug/ml
3) N-DODECANE	6.521	6740508	49.561 ug/ml
4) N-TETRADECANE	8.312	6562545	48.642 ug/ml
5) N-HEXADECANE	9.908	6283814	45.999 ug/ml
6) N-OCTADECANE	11.346	6171637	44.344 ug/ml
7) N-EICOSANE	12.652	5899691	43.028 ug/ml
8) N-DOCOSANE	13.847	5632730	42.099 ug/ml
10) N-TETRACOSANE	14.948	5499861	41.739 ug/ml
11) N-HEXACOSANE	15.970	5358537	41.936 ug/ml
12) N-OCTACOSANE	16.917	5300027	42.850 ug/ml
13) N-TRIACONTANE	17.803	5348699	44.631 ug/ml
14) N-DOTRIACONTANE	18.633	5320810	45.416 ug/ml
15) N-TETRATRIACONTANE	19.413	5266651	50.372 ug/ml
16) N-HEXATRIACONTANE	20.149	5028611	53.537 ug/ml
17) N-OCTATRIACONTANE	20.893	4699114	57.004 ug/ml
18) N-TETRACONTANE	21.821	4762631	51.460 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF072021\
Data File : FF008860.D
Signal(s) : FID2B.ch
Acq On : 20 Jul 2021 14:02
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 21 01:09:09 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF062321.M
Quant Title :
QLast Update : Thu Jun 24 08:07:11 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

