

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF072320\
 Data File : FF007209.D
 Signal(s) : FID2B.ch
 Acq On : 23 Jul 2020 10:41
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 24 01:44:08 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF071520.M
 Quant Title :
 QLast Update : Thu Jul 16 09:27:07 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.752	6739840	47.808 ug/ml
Target Compounds			
2) N-DECANE	4.465	7919083	51.483 ug/ml
3) N-DODECANE	6.492	7992641	51.457 ug/ml
4) N-TETRADECANE	8.289	7986118	51.578 ug/ml
5) N-HEXADECANE	9.892	8004822	50.773 ug/ml
6) N-OCTADECANE	11.337	8100404	50.254 ug/ml
7) N-EICOSANE	12.649	8001464	49.547 ug/ml
8) N-DOCOSANE	13.851	7869161	48.524 ug/ml
10) N-TETRACOSANE	14.957	7744147	47.287 ug/ml
11) N-HEXACOSANE	15.979	7437569	45.741 ug/ml
12) N-OCTACOSANE	16.932	7220732	44.066 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF072320\
Data File : FF007209.D
Signal(s) : FID2B.ch
Acq On : 23 Jul 2020 10:41
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 24 01:44:08 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF071520.M
Quant Title :
QLast Update : Thu Jul 16 09:27:07 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

