

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF072522\
Data File : FF011087.D
Signal(s) : FID2B.ch
Acq On : 25 Jul 2022 16:36
Operator : YP\AJ
Sample : N3814-10 10X
Misc :
ALS Vial : 65 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jul 26 02:54:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
Quant Title :
QLast Update : Fri Jul 01 01:40:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
Target Compounds			
7) N-EICOSANE	12.753	88175	0.820 ug/ml
15) N-TETRATRIACONTANE	19.452	209652	2.150 ug/ml
18) N-TETRACONTANE	21.793	84732	0.994 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF072522\
Data File : FF011087.D
Signal(s) : FID2B.ch
Acq On : 25 Jul 2022 16:36
Operator : YP\AJ
Sample : N3814-10 10X
Misc :
ALS Vial : 65 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jul 26 02:54:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF063022.M
Quant Title :
QLast Update : Fri Jul 01 01:40:42 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#7 N-EICOSANE


R.T.: 12.753 min
Delta R.T.: 0.032 min
Response: 88175
Conc: 0.82 ug/ml

Instrument :
FID_F
ClientSampleId :




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 0.000 min
Exp R.T. : 14.817 min
Response: 0
Conc: N.D.



#15 N-TETRATRIACONTANE

R.T.: 19.452 min
Delta R.T.: -0.035 min
Response: 209652
Conc: 2.15 ug/ml



#18 N-TETRACONTANE

R.T.: 21.793 min
Delta R.T.: 0.032 min
Response: 84732
Conc: 0.99 ug/ml