

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF072718\
 Data File : FF001517.D
 Signal(s) : FID2B.ch
 Acq On : 27 Jul 2018 10:57
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 27 13:54:29 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF070218.M
 Quant Title :
 QLast Update : Mon Jul 02 13:42:46 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1	N-OCTANE	124.484	111.677 E3	10.3	87	-0.02
2	N-DECANE	124.157	115.023 E3	7.4	90	-0.03
3	N-DODECANE	125.303	115.895 E3	7.5	90	-0.03
4	N-TETRADECANE	125.769	115.238 E3	8.4	90	-0.03
5	N-HEXADECANE	124.943	114.310 E3	8.5	90	-0.03
6	N-OCTADECANE	125.322	113.257 E3	9.6	89	-0.03
7	N-EICOSANE	124.391	110.714 E3	11.0	88	-0.03
8	N-DOCOSANE	123.310	108.558 E3	12.0	87	-0.03
9 S	TETRACOSANE-d50 (SURROGATE)	107.724	94.474 E3	12.3	87	-0.03
10	N-TETRACOSANE	121.908	105.555 E3	13.4	85	-0.03
11	N-HEXACOSANE	121.214	103.293 E3	14.8	84	-0.03
12	N-OCTACOSANE	120.472	102.303 E3	15.1	84	-0.03
13	N-TRIACONTANE	120.210	103.334 E3	14.0	86	-0.03
14	N-DOTRIACONTANE	117.903	104.716 E3	11.2	89	-0.03
15	N-TETRATRIACONTANE	117.532	108.238 E3	7.9	92	-0.03
16	N-HEXATRIACONTANE	113.593	108.408 E3	4.6	95	-0.03
17	N-OCTATRIACONTANE	103.255	101.415 E3	1.8	98	-0.03
18	N-TETRACONTANE	102.400	102.414 E3	-0.0	99	-0.04

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0