

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF072718\
 Data File : FF001524.D
 Signal(s) : FID2B.ch
 Acq On : 27 Jul 2018 14:50
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 27 15:12:39 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF070218.M
 Quant Title :
 QLast Update : Mon Jul 02 13:42:46 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1	N-OCTANE	124.484	114.044 E3	8.4	89	-0.02
2	N-DECANE	124.157	117.444 E3	5.4	92	-0.03
3	N-DODECANE	125.303	117.567 E3	6.2	92	-0.03
4	N-TETRADECANE	125.769	116.868 E3	7.1	91	-0.03
5	N-HEXADECANE	124.943	115.898 E3	7.2	91	-0.03
6	N-OCTADECANE	125.322	114.574 E3	8.6	90	-0.03
7	N-EICOSANE	124.391	111.817 E3	10.1	89	-0.03
8	N-DOCOSANE	123.310	109.502 E3	11.2	88	-0.03
9 S	TETRACOSANE-d50 (SURROGATE)	107.724	95.148 E3	11.7	87	-0.03
10	N-TETRACOSANE	121.908	106.306 E3	12.8	86	-0.03
11	N-HEXACOSANE	121.214	103.960 E3	14.2	85	-0.03
12	N-OCTACOSANE	120.472	102.880 E3	14.6	85	-0.03
13	N-TRIACONTANE	120.210	104.056 E3	13.4	86	-0.03
14	N-DOTRIACONTANE	117.903	105.612 E3	10.4	89	-0.03
15	N-TETRATRIACONTANE	117.532	109.228 E3	7.1	93	-0.03
16	N-HEXATRIACONTANE	113.593	109.233 E3	3.8	96	-0.03
17	N-OCTATRIACONTANE	103.255	102.418 E3	0.8	99	-0.03
18	N-TETRACONTANE	102.400	103.408 E3	-1.0	100	-0.04

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0