

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF072720\
 Data File : FF007256.D
 Signal(s) : FID2B.ch
 Acq On : 27 Jul 2020 12:43
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: Sample.e
 Quant Time: Jul 28 00:49:58 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF071520.M
 Quant Title :
 QLast Update : Thu Jul 16 09:27:07 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.755	7210207	51.145 ug/ml
Target Compounds				
2)	N-DECANE	4.466	8592224	55.859 ug/ml
3)	N-DODECANE	6.493	8656949	55.734 ug/ml
4)	N-TETRADECANE	8.291	8633238	55.757 ug/ml
5)	N-HEXADECANE	9.894	8634643	54.768 ug/ml
6)	N-OCTADECANE	11.339	8716626	54.077 ug/ml
7)	N-EICOSANE	12.652	8580215	53.131 ug/ml
8)	N-DOCOSANE	13.854	8417120	51.903 ug/ml
10)	N-TETRACOSANE	14.961	8269873	50.497 ug/ml
11)	N-HEXACOSANE	15.983	7944512	48.858 ug/ml
12)	N-OCTACOSANE	16.937	7752992	47.314 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF072720\
Data File : FF007256.D
Signal(s) : FID2B.ch
Acq On : 27 Jul 2020 12:43
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: Sample.e
Quant Time: Jul 28 00:49:58 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF071520.M
Quant Title :
QLast Update : Thu Jul 16 09:27:07 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

