

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF073118\
 Data File : FF001605.D
 Signal(s) : FID2B.ch
 Acq On : 01 Aug 2018 06:52
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 01 07:18:10 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF073018.M
 Quant Title :
 QLast Update : Mon Jul 30 13:23:50 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.340	6252602	49.762 ug/ml
Target Compounds			
2) N-DECANE	4.124	7700972	52.222 ug/ml
3) N-DODECANE	6.131	7672941	51.228 ug/ml
4) N-TETRADECANE	7.916	7581672	50.395 ug/ml
5) N-HEXADECANE	9.509	7520341	50.358 ug/ml
6) N-OCTADECANE	10.944	7455739	49.882 ug/ml
7) N-EICOSANE	12.250	7311782	49.506 ug/ml
8) N-DOCOSANE	13.444	7180312	49.529 ug/ml
10) N-TETRACOSANE	14.544	7070983	49.877 ug/ml
11) N-HEXACOSANE	15.561	7012644	50.105 ug/ml
12) N-OCTACOSANE	16.510	7006034	50.492 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF073118\
Data File : FF001605.D
Signal(s) : FID2B.ch
Acq On : 01 Aug 2018 06:52
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 01 07:18:10 2018
Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_F\Method\FF073018.M
Quant Title :
QLast Update : Mon Jul 30 13:23:50 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

