

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF080218\
 Data File : FF001631.D
 Signal(s) : FID2B.ch
 Acq On : 02 Aug 2018 20:39
 Operator : AJ\SJ
 Sample : PB111720BS
 Misc :
 ALS Vial : 65 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 03 07:35:11 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073018.M
 Quant Title :
 QLast Update : Mon Jul 30 13:23:50 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.344	1757644	13.989 ug/ml
Target Compounds			
2) N-DECANE	4.124	2080006	14.105 ug/ml
3) N-DODECANE	6.131	2108168	14.075 ug/ml
4) N-TETRADECANE	7.916	2143209	14.246 ug/ml
5) N-HEXADECANE	9.510	2104172	14.090 ug/ml
6) N-OCTADECANE	10.946	2089780	13.981 ug/ml
7) N-EICOSANE	12.252	2067075	13.996 ug/ml
8) N-DOCOSANE	13.447	2056655	14.187 ug/ml
10) N-TETRACOSANE	14.547	2119020	14.947 ug/ml
11) N-HEXACOSANE	15.566	2014312	14.392 ug/ml
12) N-OCTACOSANE	16.513	2020815	14.564 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF080218\
Data File : FF001631.D
Signal(s) : FID2B.ch
Acq On : 02 Aug 2018 20:39
Operator : AJ\SJ
Sample : PB111720BS
Misc :
ALS Vial : 65 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 03 07:35:11 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073018.M
Quant Title :
QLast Update : Mon Jul 30 13:23:50 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

