

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF080718\
 Data File : FF001724.D
 Signal(s) : FID2B.ch
 Acq On : 07 Aug 2018 17:36
 Operator : AJ\SJ
 Sample : 4400 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 08 03:53:11 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\C44-FF080418.M
 Quant Title :
 QLast Update : Mon Aug 06 12:04:39 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
13) S	TETRACOSANE-d50 (SURR...	14.338	17807000	44.175 ug/ml
Target Compounds				
1)	N-OCTANE	2.134	107809548	375.378 ug/ml
2)	N-NONANE	3.110	111289158	380.270 ug/ml
3)	N-DECANE	4.168	171804878	379.943 ug/ml
4)	N-UNDECANE	5.195	173630477	568.257 ug/ml
5)	N-DODECANE	6.170	173400445	566.628 ug/ml
6)	N-TETRADECANE	7.956	175322064	561.829 ug/ml
7)	N-HEXADECANE	9.546	145991032	463.688 ug/ml
8)	N-OCTADECANE	10.963	72404753	229.542 ug/ml
9)	N-EICOSANE	12.252	28637989	90.780 ug/ml
10)	N-TETRACOSANE	14.547	28082560	88.402 ug/ml
11)	N-OCTACOSANE	16.503	13328077	41.416 ug/ml
12)	N-DOTRIACONTANE	18.219	12890763	39.509 ug/ml
14)	N-HEXATRIACONTANE	19.741	11681682	39.601 ug/mlm
15)	N-TETRACONTANE	21.231	10281658	42.703 ug/ml
16)	N-TETRTETRACOOCTANE	23.571	7260470	45.425 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF080718\
Data File : FF001724.D
Signal(s) : FID2B.ch
Acq On : 07 Aug 2018 17:36
Operator : AJ\SJ
Sample : 4400 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 08 03:53:11 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\C44-FF080418.M
Quant Title :
QLast Update : Mon Aug 06 12:04:39 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

