

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF080918\
 Data File : FF001759.D
 Signal(s) : FID2B.ch
 Acq On : 09 Aug 2018 16:47
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 10 01:14:44 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073018.M
 Quant Title :
 QLast Update : Fri Aug 03 12:00:27 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.322	5922663	46.893 ug/ml
Target Compounds			
2) N-DECANE	4.105	6850183	46.973 ug/ml
3) N-DODECANE	6.113	7021965	47.107 ug/ml
4) N-TETRADECANE	7.899	7092268	47.056 ug/ml
5) N-HEXADECANE	9.491	7108237	47.482 ug/ml
6) N-OCTADECANE	10.926	7084575	47.003 ug/ml
7) N-EICOSANE	12.230	6951150	46.638 ug/ml
8) N-DOCOSANE	13.425	6817243	46.672 ug/ml
10) N-TETRACOSANE	14.525	6737066	47.227 ug/ml
11) N-HEXACOSANE	15.543	6419714	45.574 ug/ml
12) N-OCTACOSANE	16.490	6226078	44.403 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF080918\
Data File : FF001759.D
Signal(s) : FID2B.ch
Acq On : 09 Aug 2018 16:47
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 10 01:14:44 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073018.M
Quant Title :
QLast Update : Fri Aug 03 12:00:27 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

