

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF080918\
 Data File : FF001776.D
 Signal(s) : FID2B.ch
 Acq On : 10 Aug 2018 01:33
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 10 08:51:42 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073018.M
 Quant Title :
 QLast Update : Fri Aug 03 12:00:27 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.325	5370640	42.522 ug/mlm
Target Compounds			
2) N-DECANE	4.106	6692057	45.888 ug/ml
3) N-DODECANE	6.115	6815181	45.720 ug/ml
4) N-TETRADECANE	7.900	6814199	45.211 ug/ml
5) N-HEXADECANE	9.493	6741027	45.029 ug/ml
6) N-OCTADECANE	10.928	6632844	44.006 ug/ml
7) N-EICOSANE	12.234	6410384	43.010 ug/ml
8) N-DOCOSANE	13.428	6186830	42.356 ug/ml
10) N-TETRACOSANE	14.529	6057504	42.463 ug/ml
11) N-HEXACOSANE	15.547	5803389	41.199 ug/ml
12) N-OCTACOSANE	16.495	5703087	40.673 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF080918\
Data File : FF001776.D
Signal(s) : FID2B.ch
Acq On : 10 Aug 2018 01:33
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 10 08:51:42 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073018.M
Quant Title :
QLast Update : Fri Aug 03 12:00:27 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

