

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081023AL\
 Data File : FF013185.D
 Signal(s) : FID2B.ch
 Acq On : 10 Aug 2023 18:40
 Operator : YP\AJ
 Sample : 03572-07
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2023

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/11/2023
 Supervised By :Ankita Jodhani 08/11/2023

Integration File: autoint1.e
 Quant Time: Aug 10 22:58:54 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\Aliphatic EPH 081023.M
 Quant Title : GC Extractables
 QLast Update : Thu Aug 10 16:05:52 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	13.333	4857190	46.350 ug/ml
Spiked Amount 50.000		Recovery =	92.70%
Target Compounds			
1) T n-Nonane (C9)	3.285	405895	3.799 ug/ml
2) T n-Decane (C10)	4.539	399892	3.732 ug/ml
4) T n-Dodecane (C12)	6.702	411245	3.801 ug/ml
6) T n-Tetradecane (C14)	8.527	416034	3.854 ug/ml
7) T n-Hexadecane (C16)	10.134	417536	3.852 ug/ml
8) T n-Octadecane (C18)	11.576	436460	3.986 ug/ml
10) T n-Eicosane (C20)	12.886	422629	3.851 ug/ml
11) T n-Heneicosane (C21)	13.497	424850	3.867 ug/ml
13) T n-Docosane (C22)	14.084	423237	3.855 ug/ml
14) T n-Tetracosane (C24)	15.185	422459	3.845 ug/ml
15) T n-Hexacosane (C26)	16.206	422345	3.864 ug/ml
16) T n-Octacosane (C28)	17.156	418549	3.876 ug/ml
17) T n-Tricontane (C30)	18.042	425517	3.891 ug/ml
18) T n-Dotriacontane (C32)	18.874	439629	3.892 ug/ml
19) T n-Tetratriacontane (C34)	19.657	411063	3.870 ug/ml
20) T n-Hexatriacontane (C36)	20.395	401720	3.811 ug/mlm
21) T n-Octatriacontane (C38)	21.123	374007	3.703 ug/ml
22) T n-Tetracontane (C40)	22.021	383167	3.855 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081023AL\
Data File : FF013185.D
Signal(s) : FID2B.ch
Acq On : 10 Aug 2023 18:40
Operator : YP\AJ
Sample : 03572-07
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
LOD-MDL-WATER-01-QT3-2023

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/11/2023
Supervised By :Ankita Jodhani 08/11/2023

Integration File: autoint1.e
Quant Time: Aug 10 22:58:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\Aliphatic EPH 081023.M
Quant Title : GC Extractables
QLast Update : Thu Aug 10 16:05:52 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

