

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081120\
 Data File : FF007317.D
 Signal(s) : FID2B.ch
 Acq On : 11 Aug 2020 10:46
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Aug 11 15:27:04 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF072820.M
 Quant Title :
 QLast Update : Wed Jul 29 09:33:32 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.756	5928194	43.430 ug/ml
Target Compounds			
2) N-DECANE	4.464	7531557	53.964 ug/ml
3) N-DODECANE	6.492	7542293	53.025 ug/ml
4) N-TETRADECANE	8.291	7458993	51.258 ug/ml
5) N-HEXADECANE	9.895	7385877	49.695 ug/ml
6) N-OCTADECANE	11.340	7369789	48.062 ug/ml
7) N-EICOSANE	12.654	7174848	46.355 ug/ml
8) N-DOCOSANE	13.855	6970959	44.657 ug/ml
10) N-TETRACOSANE	14.962	6814402	43.145 ug/ml
11) N-HEXACOSANE	15.986	6563273	41.863 ug/ml
12) N-OCTACOSANE	16.939	6445830	41.070 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081120\
Data File : FF007317.D
Signal(s) : FID2B.ch
Acq On : 11 Aug 2020 10:46
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Aug 11 15:27:04 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF072820.M
Quant Title :
QLast Update : Wed Jul 29 09:33:32 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

