

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081219\
 Data File : FF005250.D
 Signal(s) : FID2B.ch
 Acq On : 12 Aug 2019 21:00
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 13 01:51:13 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073019.M
 Quant Title :
 QLast Update : Tue Jul 30 14:33:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.663	5825410	52.896 ug/ml
Target Compounds			
2) N-DECANE	4.378	7530250	52.267 ug/ml
3) N-DODECANE	6.401	7472501	51.213 ug/ml
4) N-TETRADECANE	8.196	7278074	50.565 ug/ml
5) N-HEXADECANE	9.799	7151260	50.436 ug/ml
6) N-OCTADECANE	11.244	6918649	50.529 ug/ml
7) N-EICOSANE	12.558	6849056	51.046 ug/ml
8) N-DOCOSANE	13.760	6707612	51.846 ug/ml
10) N-TETRACOSANE	14.868	6685310	53.063 ug/ml
11) N-HEXACOSANE	15.895	6708311	54.291 ug/ml
12) N-OCTACOSANE	16.849	6817341	54.576 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081219\
Data File : FF005250.D
Signal(s) : FID2B.ch
Acq On : 12 Aug 2019 21:00
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 13 01:51:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073019.M
Quant Title :
QLast Update : Tue Jul 30 14:33:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

