

rters

Instrument :
FID_F
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/15/2022
Supervised By :Ankita Jodhani 08/15/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF08122
Data File : FF011241.D
Signal(s) : FID2B.ch
Acq On : 12 Aug 2022 19:13
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF072822.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	1.852	1.772	1.935	BB	413366	5232612	93.81%	5.571%
2	4.360	4.313	4.420	BB	529229	5508484	98.75%	5.865%
3	6.530	6.481	6.592	BB	549519	5538755	99.29%	5.897%
4	8.355	8.314	8.420	BB	517760	5513461	98.84%	5.870%
5	9.959	9.911	10.017	BB	504392	5578147	100.00%	5.939%
6	11.400	11.348	11.450	BB	481165	5564838	99.76%	5.925%
7	12.708	12.655	12.760	BB	455489	5477032	98.19%	5.831%
8	13.904	13.866	13.945	VB	420703	5284624	94.74%	5.626%
9	14.802	14.751	14.833	BV	326502	4577703	82.06%	4.874%
10	15.006	14.934	15.064	BB	399152	5236143	93.87%	5.575%
11	16.025	15.972	16.067	BB	383101	5094500	91.33%	5.424%
12	16.973	16.899	17.016	BB	375861	5043248	90.41%	5.369%
13	17.859	17.792	17.895	BV	367735	5083521	91.13%	5.412%
14	18.690	18.509	18.730	BB	362811	5136454	92.08%	5.469%
15	19.472	19.261	19.509	BV	356451	5089231	91.24%	5.418%
16	20.210	20.157	20.247	BV	331396	4883331	87.54%	5.199%
17	20.914	20.867	20.955	M	294035	4916198	88.13%	5.234%
18	21.737	21.644	21.811	BV	235568	5168881	92.66%	5.503%
Sum of corrected areas:							93927163	

FF072822.M Mon Aug 15 00:41:39 2022