

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081318\
 Data File : FF001860.D
 Signal(s) : FID2B.ch
 Acq On : 13 Aug 2018 12:37
 Operator : AJ\SJ
 Sample : J4430-11MSD
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 UST-EMSD

Manual Integrations
 APPROVED

Sohil
 8/14/2018 11:36:24 AM

Integration File: autoint1.e
 Quant Time: Aug 14 02:03:15 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073018.M
 Quant Title :
 QLast Update : Fri Aug 03 12:00:27 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.314	3020117	23.912 ug/ml
Target Compounds			
1) N-OCTANE	2.040	2016797	13.905 ug/mlm
2) N-DECANE	4.097	3297085	22.608 ug/ml
3) N-DODECANE	6.105	3471190	23.287 ug/ml
4) N-TETRADECANE	7.890	3701953	24.562 ug/ml
5) N-HEXADECANE	9.483	3701874	24.728 ug/ml
6) N-OCTADECANE	10.918	3689848	24.481 ug/ml
7) N-EICOSANE	12.223	3636168	24.396 ug/ml
8) N-DOCOSANE	13.417	3581321	24.518 ug/ml
10) N-TETRACOSANE	14.516	3731968	26.161 ug/ml
11) N-HEXACOSANE	15.535	3404911	24.172 ug/ml
12) N-OCTACOSANE	16.483	3281984	23.406 ug/mlm
13) N-TRIACONTANE	17.369	3168529	22.888 ug/mlm
14) N-DOTRIACONTANE	18.201	3029967	23.225 ug/ml
15) N-TETRATRIACONTANE	18.983	2017839	16.694 ug/mlm
16) N-HEXATRIACONTANE	19.722	1057137	9.501 ug/ml
17) N-OCTATRIACONTANE	20.425	662723	6.650 ug/ml
18) N-TETRACONTANE	21.208	600951	6.032 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081318\
Data File : FF001860.D
Signal(s) : FID2B.ch
Acq On : 13 Aug 2018 12:37
Operator : AJ\SJ
Sample : J4430-11MSD
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
FID_F
Client Sampled :
UST-EMSD

Manual Integrations
APPROVED

Sohil
8/14/2018 11:36:24 AM

Integration File: autoint1.e
Quant Time: Aug 14 02:03:15 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073018.M
Quant Title :
QLast Update : Fri Aug 03 12:00:27 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

