

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081318\
 Data File : FF001864.D
 Signal(s) : FID2B.ch
 Acq On : 13 Aug 2018 14:40
 Operator : AJ\SJ
 Sample : PB111964BS
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :

Manual Integrations
 APPROVED

Sohil
 8/14/2018 11:36:28 AM

Integration File: autoint1.e
 Quant Time: Aug 14 02:04:08 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073018.M
 Quant Title :
 QLast Update : Fri Aug 03 12:00:27 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.313	2012206	15.932 ug/ml
Target Compounds			
1) N-OCTANE	2.040	2055290	14.171 ug/mlm
2) N-DECANE	4.096	2495758	17.114 ug/ml
3) N-DODECANE	6.104	2553078	17.128 ug/ml
4) N-TETRADECANE	7.889	2638690	17.507 ug/ml
5) N-HEXADECANE	9.481	2626778	17.546 ug/ml
6) N-OCTADECANE	10.917	2610394	17.319 ug/ml
7) N-EICOSANE	12.222	2568303	17.232 ug/ml
8) N-DOCOSANE	13.416	2503011	17.136 ug/ml
10) N-TETRACOSANE	14.516	2596485	18.202 ug/ml
11) N-HEXACOSANE	15.534	2353498	16.708 ug/ml
12) N-OCTACOSANE	16.482	2328262	16.605 ug/ml
13) N-TRIACONTANE	17.367	2299560	16.611 ug/ml
14) N-DOTRIACONTANE	18.199	2175063	16.672 ug/ml
15) N-TETRATRIACONTANE	18.981	1523060	12.600 ug/ml
16) N-HEXATRIACONTANE	19.720	771081	6.930 ug/ml
17) N-OCTATRIACONTANE	20.423	480688	4.823 ug/ml
18) N-TETRACONTANE	21.207	424477	4.261 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081318\
Data File : FF001864.D
Signal(s) : FID2B.ch
Acq On : 13 Aug 2018 14:40
Operator : AJ\SJ
Sample : PB111964BS
Misc :
ALS Vial : 61 Sample Multiplier: 1

Instrument :
FID_F
ClientSampled :

Manual Integrations
APPROVED

Sohil
8/14/2018 11:36:28 AM

Integration File: autoint1.e
Quant Time: Aug 14 02:04:08 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073018.M
Quant Title :
QLast Update : Fri Aug 03 12:00:27 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

