

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081319\
 Data File : FF005260.D
 Signal(s) : FID2B.ch
 Acq On : 13 Aug 2019 09:45
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 13 22:33:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073019.M
 Quant Title :
 QLast Update : Tue Jul 30 14:33:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.671	5530377	50.217 ug/ml
Target Compounds			
2) N-DECANE	4.381	7084036	49.170 ug/ml
3) N-DODECANE	6.405	7094044	48.619 ug/ml
4) N-TETRADECANE	8.201	6945654	48.256 ug/ml
5) N-HEXADECANE	9.804	6850365	48.314 ug/ml
6) N-OCTADECANE	11.250	6637342	48.475 ug/ml
7) N-EICOSANE	12.564	6565918	48.936 ug/ml
8) N-DOCOSANE	13.768	6412716	49.566 ug/ml
10) N-TETRACOSANE	14.876	6355950	50.449 ug/ml
11) N-HEXACOSANE	15.903	6346290	51.361 ug/ml
12) N-OCTACOSANE	16.858	6434123	51.509 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081319\
Data File : FF005260.D
Signal(s) : FID2B.ch
Acq On : 13 Aug 2019 09:45
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 13 22:33:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073019.M
Quant Title :
QLast Update : Tue Jul 30 14:33:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

