

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081420\
 Data File : FF007384.D
 Signal(s) : FID2B.ch
 Acq On : 14 Aug 2020 14:50
 Operator : DD\AJ
 Sample : L3678-04MS
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 17 01:49:20 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF072820.M
 Quant Title :
 QLast Update : Wed Jul 29 09:33:32 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.754	3961937	29.025 ug/ml
Target Compounds			
1) N-OCTANE	2.344	1285617	9.477 ug/ml
2) N-DECANE	4.460	2232553	15.996 ug/ml
3) N-DODECANE	6.486	2416140	16.986 ug/ml
4) N-TETRADECANE	8.284	2595581	17.837 ug/ml
5) N-HEXADECANE	9.889	3091162	20.798 ug/ml
6) N-OCTADECANE	11.334	2440468	15.916 ug/mlm
7) N-EICOSANE	12.649	3685631	23.812 ug/ml
8) N-DOCOSANE	13.850	3287484	21.060 ug/ml
10) N-TETRACOSANE	14.961	11466584	72.600 ug/ml
11) N-HEXACOSANE	15.985	2890562	18.437 ug/ml
12) N-OCTACOSANE	16.938	2462330	15.689 ug/ml
13) N-TRIACONTANE	17.831	3864509	24.641 ug/ml
14) N-DOTRIACONTANE	18.641	2099012	14.167 ug/ml
15) N-TETRATRIACONTANE	19.454	2187274	18.085 ug/ml
16) N-HEXATRIACONTANE	20.200	2625342	28.383 ug/ml
17) N-OCTATRIACONTANE	20.965	2190966	30.846 ug/mlm
18) N-TETRACONTANE	21.918	2968071	44.378 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081420\
Data File : FF007384.D
Signal(s) : FID2B.ch
Acq On : 14 Aug 2020 14:50
Operator : DD\AJ
Sample : L3678-04MS
Misc :
ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 17 01:49:20 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF072820.M
Quant Title :
QLast Update : Wed Jul 29 09:33:32 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

